



C-VUE Advanced® HydraVUE™ (efofilcon A) SOFT (hydrophilic) SILICONE HYDROGEL CONTACT LENS for DAILY WEAR

®/TM are trademarks of Bausch & Lomb Incorporated or its affiliates.
© 2019 Bausch & Lomb Incorporated or its affiliates.

Rev. 2019-06 8170203

Rx ONLY CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE
TO SALE BY OR ON THE ORDER OF A LICENSED
PRACTITIONER.

IMPORTANT

Please read carefully and keep this information for future use.

DESCRIPTION

The non-ionic lens material, efofilcon A, is a daily wear silicone hydrogel contact lens that is not surface treated and is characterized by a high-water content. The lens material is composed of silicone monomers cross linked with other monomers and optionally incorporates D&C Green 6 as an integrated handling tint. The lenses are lathe-cut and made to order for individual prescriptions. The material consists of 26% efofilcon A and 74% water by weight when immersed in saline solution.

In the hydrated state, the lens conforms to the curvature of the eye covering the cornea and extends slightly beyond the limbus forming a colorless, transparent optical surface. The hydrophilic properties of the lens require that it be maintained in a fully hydrated state in a solution compatible with the eye. If the lens dries out, it will become hard and appear somewhat warped; however, it will return to its proper configuration when completely rehydrated in the proper storage solution.

The physical/optical properties of the lenses are:

Refractive Index:	1.38 (hydrated)
Light Transmission:	Greater than 97%
Surface Character:	Hydrophilic
Water Content:	74%
Specific Gravity:	1.048 (hydrated)
Oxygen Permeability:	59.8 x 10 ⁻¹¹ (cm ² /sec) (ml O ₂ /ml x hPa @ 35°C), (revised Fatt method).

The hydrophilic characteristics allow aqueous solutions to enter the lens and in its fully hydrated state, the lens is approximately 74% water by weight. The lenses are manufactured in Single Vision, Toric, Multifocal and Toric Multifocal configurations with the following features and properties:

Diameter:	12.0mm to 16.00mm
Center Thickness:	0.01mm to 0.50mm
Base Curve:	8.0mm to 9.5mm
Power:	-20.00D to +20.00D in 0.25 steps
Cylinder Power:	
Toric:	-0.25D to -10.00D
Toric Multifocal:	-0.25D to -4.00D
Add Power (Multifocal):	+0.50D to +4.00D

ACTIONS

In its hydrated state, the C-VUE Advanced® HydraVUE™ (efofilcon A) Soft (hydrophilic) Silicone Hydrogel Contact Lenses for daily wear, when placed on the cornea, acts as a refracting medium to focus light rays on the retina.

INDICATIONS

The C-VUE Advanced® HydraVUE™ (efofilcon A) Single Vision Soft (hydrophilic) Silicone Hydrogel Contact Lenses for **daily wear** are indicated for the correction of ametropia (myopia and hyperopia) in aphakic and non-aphakic persons with non-diseased eyes. The lens may be worn by persons who exhibit refractive astigmatism of 0.75 diopters or less where the astigmatism does not interfere with visual acuity.

BAUSCH + LOMB Unilens™

The C-VUE Advanced® HydraVUE™ (efofilcon A) Toric Soft (hydrophilic) Silicone Hydrogel Contact Lenses for **daily wear** are indicated for the correction of ametropia (myopia and hyperopia) in aphakic and non-aphakic persons with non-diseased eyes and/or possess refractive astigmatism not exceeding 10 diopters.

The C-VUE Advanced® HydraVUE™ (efofilcon A) Multifocal Soft (hydrophilic) Silicone Hydrogel Contact Lenses for **daily wear** are indicated for the correction of ametropia (myopia and hyperopia) in aphakic and non-aphakic persons with non-diseased eyes and/or possess refractive astigmatism not exceeding 0.75 diopters and are presbyopic requiring add power of up to +4.00 diopters.

The C-VUE Advanced® HydraVUE™ (efofilcon A) Toric Multifocal Soft (hydrophilic) Silicone Hydrogel Contact Lenses for **daily wear** are indicated for the correction of ametropia (myopia and hyperopia) in aphakic and non-aphakic persons with non-diseased eyes and/or possess refractive astigmatism not exceeding 4 diopters and are presbyopic requiring add power of up to +4.00 diopters.

Eye care practitioners may prescribe the lenses for frequent/planned replacement wear, with cleaning, disinfection, and scheduled replacement. When prescribed for frequent/planned replacement wear, the lens may be cleaned and disinfected using a chemical (not heat) lens care system or hydrogen peroxide disinfection systems.

CONTRAINDICATIONS (REASONS NOT TO USE)

DO NOT USE the C-VUE Advanced® HydraVUE™ (efofilcon A) Soft Silicone Hydrogel Contact Lenses when any of the following conditions are present:

- Acute and subacute inflammation or infection of the anterior chamber of the eye.
- Any eye disease, injury, or abnormality that affects the cornea, conjunctiva, or eyelids.
- Severe insufficiency of lacrimal secretion (dry eyes).
- Corneal hypoesthesia (reduced corneal sensitivity), if non-aphakic.
- Any systemic disease that may affect the eye or be exaggerated by wearing contact lenses.
- Allergic reactions of ocular surfaces or adnexa that may be induced or exaggerated by wearing contact lenses or use of contact lens solutions.
- Allergy to any ingredient, such as mercury or thimerosal, in a solution which is to be used to care for the C-VUE Advanced® HydraVUE™ (efofilcon A) Soft Silicone Hydrogel Contact Lenses.
- Any active corneal infection (bacterial, fungal, or viral).
- If eyes become red or irritated.

WARNINGS

Patients should be advised of the following warnings pertaining to contact lens wear:

- **Problems with contact lenses and lens care products could result in serious injury to the eye.** It is essential that patients follow their eye care practitioner's direction and all labeling instructions for proper use of lenses and lens care products, including the lens case. Eye problems, including corneal ulcers, can develop rapidly and lead to **loss of vision**.
- Daily wear lenses are not indicated for overnight wear, and **patients should be instructed not to wear lenses while sleeping**. Clinical studies have shown that the risk of serious adverse reactions is increased when daily wear lenses are worn overnight.
- Studies have shown that contact lens wearers who are smokers have a higher incidence of adverse reactions than nonsmokers.
- If a patient experiences eye discomfort, excessive tearing, vision changes, or redness of the eye, the patient should be instructed to **immediately remove the lenses** and promptly contact his or her eye care practitioner.

PRECAUTIONS

Special Precautions for Eye Care Practitioners

- Due to the small number of patients enrolled in clinical investigation of lenses, all refractive powers, design configurations, or lens parameters available in the lens material were not evaluated in significant numbers. Consequently, when selecting an appropriate lens design and parameters, the eye care practitioner should consider all characteristics of the lens that can affect lens performance and ocular health, including oxygen permeability, wettability, central and peripheral thickness, and optic zone diameter.
- The potential impact of these factors on the patient's ocular health must be carefully weighed against the patient's need for refractive correction; therefore, the continuing ocular health of the patient and lens performance on the eye should be carefully monitored by the prescribing eye care practitioner.
- Patients who wear aspheric contact lenses to correct presbyopia may not achieve the best corrected visual acuity for either far or near vision. Visual requirements vary with the individual and should be considered when selecting the most appropriate type of lens for each patient.
- Aphakic patients should not be fitted with C-VUE Advanced® HydraVUE™ (efofilcon A) Soft Silicone Hydrogel Contact Lenses until the determination is made that the eye has healed completely.
- Fluorescein, a yellow dye, should not be used while the lenses are on the eyes. The lenses absorb this dye and become discolored. Whenever fluorescein is used in eyes, the eyes should be flushed with a sterile saline solution that is recommended for in-eye use.
- Before leaving the eye care practitioner's office, the patient should be able to promptly remove the lens or should have someone else available who can remove the lens for him or her.
- Eye care practitioners should instruct the patient to **remove the lens immediately** if the eye becomes red or irritated.

Eye care practitioners should carefully instruct patients about the following lens care regimen and safety precautions:

- Different solutions cannot always be used together, and not all solutions are safe for use with all lenses. Use only recommended solutions that are fresh and sterile.
- Never use solutions recommended for conventional hard contact lenses only.
- Always use **fresh, unexpired** lens care solutions.
- Always follow directions in the package insert for the use of contact lens solutions.
- Use **only chemical (not heat) lens care systems** labeled for use with soft contact lenses.
- Sterile unpreserved solutions, when used, should be discarded after the time specified in the labeling directions.
- Do not use saliva or anything other than the recommended solution for lubricating or rewetting lenses.
- Tap water, distilled water, or homemade saline should not be used as a substitute for any component in the lens care regimen since they have been associated with an *Acanthamoeba* keratitis infection.
- Always keep the lenses completely immersed in the recommended storage solution when the lenses are not being worn (stored). Prolonged periods of drying will damage the lens. Follow the lens care directions in CARE FOR A DRIED OUT (DEHYDRATED) LENS in the Patient Information Booklet if the lens surface becomes dried out.
- If the lens sticks (stops moving) on the eye, follow the recommended directions in CARE FOR A STICKING (NON-MOVING) LENS. The lens should move freely on the eye for the continued health of the eye. If non-movement of the lens continues, the patient should be instructed to **immediately remove the lenses** and consult his or her eye care practitioner.
- Always wash and rinse hands before handling lenses. Do not get cosmetics, lotions, soaps, creams, deodorants, or sprays in the eyes or on the lenses. It is best to put on lenses before putting on makeup. Water-based cosmetics are less likely to damage lenses than oil-based products.
- Do not touch contact lenses with the fingers or hands if the hands are not free of foreign materials, as microscopic scratches on the lens may occur, causing distorted vision and/or injury to the eye.
- Carefully follow the handling, insertion, removal, cleaning, disinfecting, storing, and wearing instructions in the Patient Information Booklet for the C-VUE Advanced® HydraVUE™ (efofilcon A) Soft Silicone Hydrogel Contact Lenses and those prescribed by the eye care practitioner.
- Never wear lenses beyond the period recommended by the eye care practitioner.
- If aerosol products such as hair spray are used while wearing the lenses, exercise caution and keep eyes closed until the spray has settled.
- Always handle lenses carefully and avoid dropping them.
- Avoid all harmful or irritating vapors and fumes while wearing lenses.
- Ask the eye care practitioner about wearing lenses during sporting and water related activities. Exposure to water while wearing contact lenses in activities such as swimming, water skiing, and hot tubs may increase the risk of ocular infection including, but not limited to, *Acanthamoeba* keratitis.
- Inform the doctor (health care practitioner) about being a contact lens wearer.
- Never use tweezers or other tools to remove lenses from the lens container unless specifically indicated for that use. Pour the lens into the hand.
- Do not touch the lens with fingernails.
- Always discard lenses worn on a frequent/planned replacement schedule after the recommended wearing schedule prescribed by the eye care practitioner.
- Always contact the eye care practitioner before using any medicine in the eyes.

- Certain medications, such as antihistamines, decongestants, diuretics, muscle relaxants, tranquilizers, and those for motion sickness may cause dryness of the eye, increased lens awareness, or blurred vision. Should such conditions exist, proper remedial measures should be prescribed. Depending on the severity, this could include the use of lubricating drops that are indicated for use with soft contact lenses or temporary discontinuance of contact lens wear while such medication is being used.
- Oral contraceptive users could develop visual changes or changes in lens tolerance when using contact lenses. Patients should be cautioned accordingly.
- Always inform the employer of being a contact lens wearer. Some jobs may require use of eye protection equipment or may require that the patient do not wear contact lenses.
- As with any contact lens, follow-up visits are necessary to assure the continuing health of the patient's eyes. The patient should be instructed as to a recommended follow-up schedule.

ADVERSE REACTIONS (PROBLEMS AND WHAT TO DO)
The patient should be informed that the following problems may occur:

- Eyes stinging, burning, itching (irritation), or other eye pain
- Comfort is less than when the lens was first placed on the eye
- Abnormal feeling of something in the eye (foreign body, scratched area)
- Excessive watering (tearing) of the eyes
- Unusual eye secretions
- Redness of the eyes
- Reduced sharpness of vision (poor visual acuity)
- Blurred vision, rainbows, or halos around objects
- Sensitivity to light (photophobia)
- Dry eyes

If the patient notices any of the above, he or she should be instructed to:

- **Immediately remove the lenses.**
- If the discomfort or problem stops, then look closely at the lens. If the lens is in any way damaged, **DO NOT put the lens back on the eye.** Place the lens in the storage case and contact the eye care practitioner. If the lens has dirt, an eyelash, or other foreign body on it, or the problem stops, and the lens appears undamaged, the patient should thoroughly clean, rinse, and disinfect the lens; then reinsert them. After reinsertion, if the problem continues, the patient should **immediately remove the lenses and consult the eye care practitioner.**

If the above symptoms continue after removal of the lens, or upon reinsertion of a lens, or upon insertion of a new lens, the patient should **immediately remove the lenses and contact his or her eye care practitioner** or physician, who must determine the need for examination, treatment, or referral without delay. (See Important Treatment Information for Adverse Reactions.) A serious condition such as infection, corneal ulcer, corneal vascularization, or iritis may be present, and may progress rapidly. Less serious reactions such as abrasions, epithelial staining or bacterial conjunctivitis must be managed and treated carefully to avoid more serious complications.

Important Treatment Information for Adverse Reactions
Sight-threatening ocular complications associated with contact lens wear can develop rapidly, and therefore early recognition and treatment of problems are critical. Infectious corneal ulceration is one of the most serious potential complications and may be ambiguous in its early stage. Signs and symptoms of infectious corneal ulceration include discomfort, pain, inflammation, purulent discharge, sensitivity to light, cells and flare, and corneal infiltrates.

Initial symptoms of a minor abrasion and an early infected ulcer are sometimes similar. Accordingly, such epithelial defect, if not treated properly, may develop into an infected ulcer. In order to prevent serious progression of these conditions, a patient presenting symptoms of abrasions or early ulcers should be evaluated as a potential medical emergency, treated accordingly, and be referred to a corneal specialist when appropriate. Standard therapy for corneal abrasions such as eye patching or the use of steroids or steroid/antibiotic combination may exacerbate the condition. If the patient is wearing a contact lens on the affected eye when examined, the lens should be removed immediately, and the lens and lens care products retained for analysis and culturing.

FITTING
Conventional methods of fitting contact lenses do apply to the C-VUE Advanced® HydraVUE™ (efrofilcon A) Soft Silicone Hydrogel Contact Lenses. For a detailed description of the fitting techniques, refer to the C-VUE Advanced® HydraVUE™ (efrofilcon A) Soft Silicone Hydrogel Contact Lenses Professional Fitting Guide, copies of which are available from: Unilens, 21 Park Place Blvd. North, Clearwater, FL 33759 USA; Telephone: 1-800-253-3669.

WEARING SCHEDULE
The wearing and replacement schedules should be determined by the eye care practitioner. Regular checkups, as determined by the eye care practitioner, are extremely important.

Daily Wear
There may be a tendency for the daily wear patient to overwear the lenses initially. Therefore, the importance of adhering to a proper, initial daily wearing schedule should be stressed to these patients. The wearing schedule should be determined by the eye care practitioner. The wearing schedule chosen by the eye care practitioner should be provided to the patient.

Frequent/Planned Replacement Wear
When removed between replacement periods, lenses must be cleaned and disinfected before reinsertion, or be discarded and replaced with a new lens.

STUDIES HAVE NOT BEEN COMPLETED TO SHOW THAT THE C-VUE Advanced® HydraVUE™ (EFROFILCON A) SOFT SILICONE HYDROGEL CONTACT LENSES ARE SAFE TO WEAR DURING SLEEP.

HANDLING OF LENS
Patient Lens Care Directions
When lenses are dispensed, the patient should be provided with appropriate and adequate instructions and warnings for lens care handling. The eye care practitioner should recommend appropriate and adequate procedures and products for each individual patient in accordance with the particular lens wearing schedule and care system selected by the practitioner; the specific instructions for such products and the particular characteristics of the patient.

For complete information concerning the care, cleaning, and disinfection of contact lenses, refer to the C-VUE Advanced® HydraVUE™ (efrofilcon A) Soft Silicone Hydrogel Contact Lens for Daily Wear Patient Information Booklet.

Soaking and Storing Lenses
Instruction for Use:
Use only fresh contact lens disinfecting solution each time you soak (store) lenses.

WARNING:
Do not re-use or "top-off" old solution left in lens case since solution re-use reduces effective lens disinfection and could lead to severe infection, vision loss or blindness. "Topping-off" is the addition of fresh solution to solution that has been sitting in the case.

Rub and Rinse Time
Instruction for Use:
Follow the complete recommended lens rubbing and rinsing times in the labeling of the solution used for cleaning, disinfecting, and soaking lenses to adequately disinfect lenses and reduce the risk of contact lens infection.

WARNING:
Rub and rinse lenses for the recommended amount of time to help prevent serious eye infections. **Never use water,** saline solution, or rewetting drops to disinfect lenses. These solutions will not disinfect lenses. Not using the recommended disinfectant can lead to severe infection, vision loss or blindness.

Lens Case Care
Instruction for Use:
Clean contact lens cases with digital rubbing with fresh, sterile disinfecting solutions/contact lens cleaner. **Never use water.** Cleaning should be followed by rinsing with fresh, sterile disinfecting solutions (**never use water**) and wiping the lens cases with fresh, clean tissue is recommended. Air-drying or recapping the lens case lids after use without any additional cleaning methods should be avoided. If air-drying, be sure that no residual solution remains in the case before allowing it to air-dry. Replace the lens case according to the directions given by your eye care practitioner or the labeling that came with your case. Contact lens cases can be a source of bacterial growth.

WARNING:
Do not store lenses or rinse lens case with water or any non-sterile solution. Only use fresh solution so you do not contaminate lenses or lens case. Use of non-sterile solution can lead to severe infection, vision loss or blindness.

Water Activity
Instruction for Use:
Do not expose contact lenses to water while wearing them.

WARNING:
Water can harbor microorganisms that can lead to severe infection, vision loss or blindness. If your lenses have been submersed in water when swimming in pools, lakes, or oceans, discard them and replace them with a new pair. Ask your eye care practitioner for recommendations about wearing lenses during any activity involving water.

Discard Date on Solution Bottle
Instruction for Use:
Discard any remaining solution after the recommended time period indicated on the bottle of solution used for disinfecting and soaking contact lenses.

WARNING:
Using solution beyond the discard date could result in contamination of the solution and can lead to severe infection, vision loss or blindness.

CARE FOR A STICKING (NON-MOVING) LENS
If the lens sticks (stops moving), the patient should be instructed to use a lubricating or rewetting solution in their eye. The patient should be instructed to **not** use plain water or anything other than the recommended solutions. The patient should be instructed to contact the eye care practitioner if the lens does not begin to move upon blinking after several applications of the solution or drops, and to **not** attempt to remove the lens except on the advice of the eye care practitioner.







STORAGE
C-VUE Advanced® HydraVUE™ (efrofilcon A) Soft Silicone Hydrogel Contact Lenses must be stored only in the recommended solutions. If left exposed to the air, the lens will dehydrate. If lens dehydrates, follow the lens care direction in the CARE FOR A DRIED OUT (DEHYDRATED) LENS section of the Patient Information Booklet.

LENS CARE PRODUCTS
The eye care practitioner should recommend a care system that is appropriate for C-VUE Advanced® HydraVUE™ (efrofilcon A) Soft Silicone Hydrogel Contact Lenses. Each lens care product contains specific directions for use and important safety information, which should be read and carefully followed. Multi-purpose solutions are the preferred choice for use. If using hydrogen peroxide solutions, exposure to peroxide should be limited by using a disc-based system.

EMERGENCIES
The patient should be informed that if chemicals of any kind (household products, gardening solutions, laboratory chemicals, etc.) are splashed into the eyes, the patient should: **FLUSH EYES IMMEDIATELY WITH TAP WATER AND THEN REMOVE LENSES PROMPTLY. CONTACT THE EYE CARE PRACTITIONER OR VISIT A HOSPITAL EMERGENCY ROOM WITHOUT DELAY.**

REPORTING OF ADVERSE REACTIONS
All serious adverse experiences and adverse reactions observed in patients wearing C-VUE Advanced® HydraVUE™ (efrofilcon A) Soft Silicone Hydrogel Contact Lenses or experienced with the lenses, should be reported to: Unilens, 21 Park Place Blvd. North, Clearwater, FL 33759 USA; Telephone: 1-800-253-3669.

HOW SUPPLIED
Each lens is supplied sterile in a glass vial containing saline solution. The vial labeling is printed with appropriate lot numbering, expiration dating and lens parameter identification.

SYMBOLS USED ON LABELING	
Symbol	Description
	Manufacturer
	Batch Code
	Prescription Only (USA)
	Use-By Date
	Caution
	Sterilized Using Steam
CYL	Cylinder Power
AX	Cylinder Axis
BC	Base Curve
ADD	Add Power
DIA	Diameter
PWR	Power

 Unilens
21 Park Place Blvd. North
Clearwater, FL 33759 USA
1-800-253-3669

